



Certificate of Compliance

Certificate: 80022898

Master Contract: 237041

Project: 80022898

Date Issued: March 29, 2021

Issued To: GfG Instrumentation, Inc.
1194 Oak Valley Dr, Ste 20
Ann Arbor, Michigan, 48108
United States

Attention: Tony Poley

Semyon Baum
Issued by: Semyon Baum



PRODUCTS

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT Certified to US Standards

The GMA200-MT is a gas measuring controller for DIN Rail mounting with 16 analog signal inputs. 24VDC, 5 W. Relay contacts 250Vac or 30 Vdc @ 3 A.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12
+ Amd 1 - 18

Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 1: General requirements

UL 61010-1
Third Edition

Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 1: General requirements

MARKINGS

The manufacturer is required to apply the following markings:

- Products shall be marked with the markings specified by the particular product standard.



Certificate:
Project: 80022898

Master Contract: 237041
Date Issued: March 29, 2021

- Products certified for Canada shall have all Caution and Warning markings in both English and French.

Additional bilingual markings not covered by the product standard(s) may be required by the Authorities Having Jurisdiction. It is the responsibility of the manufacturer to provide and apply these additional markings, where applicable, in accordance with the requirements of those authorities.

The products listed are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US (indicating that products have been manufactured to the requirements of both Canadian and U.S. Standards) or with adjacent indicator 'US' for US only or without either indicator for Canada only.

The following markings appear on the product:

1. Submitter's identification (company name and/or file number and/or registered tradename);
2. Electrical rating;
3. Date of manufacture: Month and year of manufacture or date code. If a serial number is used instead of date of manufacture, a record of serial numbers shall be kept traceable to date of manufacture. (Not related to date of sale);
4. The products listed are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US, or with adjacent indicator 'US' for US only, or without either indicator for Canada only.



or

or



Supplement to Certificate of Compliance

Certificate: 80022898

Master Contract: 237041

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	2021-03-29	Description
80022898		Original CAN-US certification evaluation of the model GMA 200-MT Controller, fixed gas detection control unit designed for connection to multiple gas sensing modules ('transmitters'). The evaluation is for ordinary locations approval to 61010-1